

ALS Group, USA

Account # 140

USEPA Lab ID WA01276

NPDES ID #

Deb Callahan
1317 South 13th Avenue
Kelso WA 98626

Study # QTA

Open Date 01/18/2018

Study Type External PT

Close Date 01/30/2018

NELAC #	Component	Method Code	Method Description	Reported Value	AV or StudyMean	Assigned Value	Acceptance Limits Low	Acceptance Limits High	Performance Evaluation	Analysis Date
---------	-----------	-------------	--------------------	----------------	-----------------	----------------	-----------------------	------------------------	------------------------	---------------

Part#	Lot#	Water PT PFOA Mix	14 Components				Invoice#	Units ug/L		
n/a	EtFOSAA (cas# 2991-50-6)	8676	EPA 537.1	0.147	0.160	.16	.096	.224	ACCEPT.	01/24/2018
n/a	MeFOSAA (cas# 2355-31-9)	8676	EPA 537.1	0.280	0.310	.31	.186	.434	ACCEPT.	01/24/2018
6911	PFBS (cas# 375-73-5)	8676	EPA 537.1	0.179	0.188	.188	.113	.263	ACCEPT.	01/24/2018
6905	PFDA (cas# 335-76-2)	8676	EPA 537.1	0.368	0.386	.386	.232	.54	ACCEPT.	01/24/2018
6903	PFDaA (cas# 307-55-1)	8676	EPA 537.1	0.145	0.165	.165	.099	.231	ACCEPT.	01/24/2018
6908	PFHpA (cas# 375-85-9)	8676	EPA 537.1	0.234	0.256	.256	.154	.358	ACCEPT.	01/24/2018
n/a	br-PFHxS (cas# 355-46-4)	8676	EPA 537.1	0.200	0.213	.213	.128	.298	ACCEPT.	01/24/2018
6913	PFHxA (cas# 307-24-4)	8676	EPA 537.1	0.165	0.151	.151	.091	.211	ACCEPT.	01/24/2018
6906	PFNA (cas# 375-95-1)	8676	EPA 537.1	0.130	0.125	.125	.075	.175	ACCEPT.	01/24/2018
6909	br-PFOS (cas# 1763-23-1)	8676	EPA 537.1	0.372	0.446	.446	.268	.624	ACCEPT.	01/24/2018
6912	PFOA (cas# 335-67-1)	8676	EPA 537.1	0.132	0.136	.136	.082	.19	ACCEPT.	01/24/2018
6902	PFTeDA (cas# 376-06-7)	8676	EPA 537.1	0.116	0.120	.12	.072	.168	ACCEPT.	01/24/2018
9563	PFTTrDA (cas# 72629-94-8)	8676	EPA 537.1	0.131	0.160	.16	.096	.224	ACCEPT.	01/24/2018
6904	PFUnA (cas# 2058-94-8)	8676	EPA 537.1	0.204	0.225	.225	.135	.315	ACCEPT.	01/24/2018

ABSOLUTE STANDARDS, INC., ISO 9001 Registered, (NSF) • PO BOX 5585, HAMDEN, CT 06518, PHONE (203) 281-2917, FAX (203) 281-2922 (203) 281-2922

[This Form: Performance Evaluation Report Form, Rev:5, Date Issued:11162010] [This Report: 140 WS 013018.pdf, Page 1 of 1 Printed: 1/30/2018,10:27:27 AM]

Samples were prepared and scored according to the principles outlined in the "The TNI Standard EL-V3-2009" and the current Fields of Proficiency Testing Tables, FoPTs. All components are formulated and verified under Absolutes' NELAC scope (ANAB Accreditation ISO 17025, 17043 (Cert.# AP-1543) , Guide 34-35) as shown in blue font.

This report may be used in whole or in part by the participant. All results are confidential but limited to accreditation body or other participant requests!